



CALIFORNIA DEVICE RECALL INFORMATION SHEET

Imperative Care Inc. Recalls ZOOM 71 Reperfusion Catheter Due to Risk of Breaks During Use

Recall Date	Product Description	Recalling Firm	Recall Reason
08/13/2021	Zoom 71 Reperfusion Catheter 3118 units Aspiration Catheter 137 cm length 0.071” Inner Diameter	Imperative Care Inc, Campbell, California	Distal tip can break during use

Recall Class	Product Identification	Distribution	Affected Dates
I	Lot numbers: V2014902, F2026201, F2026802, F2028801, F2028901, F2029501, F2029701, F2030801, F2031101, F2035301, F2032502, F2100802, F2101301, F2104102, F2108101, F2108801, F2109601, F2109701, F2111101, F2110601, F2112301, F2113004, F2112001	323 units within California	September 2020 – August 2021

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#)

