



Food and Drug Branch – Device Recalls

CALIFORNIA DEVICE RECALL INFORMATION SHEET

iiiIntegra Lifesciences Recalls Cerelink Extension Cable For Design Changes

Recall Date	Product Description	Recalling Firm	Recall Reason
12/20/2023	Cerelink ICP Extension Cable (Extension Cable) Model No. 826845 Used with Codman CereLink Intracranial Pressure (ICP) Monitor, Model No. 826820/826820P. Cable is also provided with monitor as part of the same SKU.	INTEGRA LIFESCIENCES CORP. Princeton, New Jersey	Firm is expanding the existing recall RES 90457 for the Cerelink, initiated in June 2022, to include all CereLink ICP Extension Cables. The updated corrective action includes a design change to the extension cable to address the "out of range readings" issue.

Recall Class	Product Identification	Distribution	Affected Dates
I	Cerelink ICP Extension Cable UDI-DI 10381780520665 All lots distributed	4941 Units Nationwide including California	October 2023 and prior

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#)

