



Food and Drug Branch – Device Recalls

CALIFORNIA DEVICE RECALL INFORMATION SHEET

**All Ultrasound Gels and Lotions Manufactured by Eco-Med Pharmaceutical, Inc.
 Recalled Due to Risk of Bacteria Contamination**

Recall Date	Product Description	Recalling Firm	Recall Reason
11/3/2021	IPL Coupling Gel	Lumenis Inc. Salt Lake City, Utah	Ultrasound Gels and lotions manufactured by Eco-Med Pharmaceutical may contain bacterial contamination

Recall Class	Product Identification	Distribution	Affected Dates
I	IPL Coupling Gel 1 L (AX1009013) 12 pack 0.25 L (AX1009018) 6 pack of 1 L (AX1009019 or AX1009063) 0.25 L (AX1009012) Ecogel 200	11,064 units Nationwide including California	Last 3 years

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#)

