



Food and Drug Branch – Device Recalls

CALIFORNIA DEVICE RECALL INFORMATION SHEET

Med-Mizer recalls MR600 and STS500 Patient Lifts

| Recall Date | Product Description | Recalling Firm | Recall Reason |
|-------------|--|--|--|
| 3/8/2023 | Med-Riser MR600 Patient Lift | Med-Mizer Inc. Batesville, Indiana | Risk of boom pivot failing due to the nut securing the boom becoming loose and the bolt to be unsecured to the flange on the side. |
| 3/8/2023 | Sit to Stand STS500 Patient Lift | Med-Mizer Inc. Batesville, Indiana | Risk of boom pivot failing |

| Recall Class | Product Identification | Distribution | Affected Dates |
|--------------|---|-----------------------|-------------------------|
| II | Med-Riser MR600 UDI-DI: 00852195007308; Serial No.: MR600000 - MR600240 | 8 Units in California | February 2023 and prior |
| II | Sit to Stand STS500 UDI-DI: 00852195007353; Serial No.: STS500000 - STS500036 | 1 Unit in California | February 2023 and prior |

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#)

CDPH Food and Drug Branch
 MS 7602 • P.O. Box 997435 • Sacramento, CA 95899-7435
 (916) 650-6500 • (916) 650-6650 FAX
 Internet Address: www.cdph.ca.gov

