



Food and Drug Branch – Device Recalls

CALIFORNIA DEVICE RECALL INFORMATION SHEET

Medline Industries Recalls Procedure Packs For Water Damage

Recall Date	Product Description	Recalling Firm	Recall Reason
6/8/2023	Procedure Packs (1) Medline Sterile Arthroscopy Pack, REF: DYNJ64242B; (2) Medline Sterile Shoulder Pack, REF: DYNJ64250A; and (3) Medline Sterile Upper Extremity Pack, REF: DYNJ64253B.	MEDLINE INDUSTRIES, LP - NORTHFIELD Northfield, Illinois	The kits were damaged by water.

Recall Class	Product Identification	Distribution	Affected Dates
II	Procedure Packs (1) Arthroscopy Pack - Lot #22HMD098, Exp. 2024-04-30, UDI/DI (01)10193489956405; (2) Shoulder Pack -Lot #22JMA673, Exp. 2024-01-31, GTIN (01)10193489429817; and (3) Upper Extremity Pack - Lot #HMF900, Exp. 2024-01-31, UDI/DI (01)10193489968453.	342 Units in California	April 2023 and Prior

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#)

CDPH Food and Drug Branch
 MS 7602 • P.O. Box 997435 • Sacramento, CA 95899-7435
 (916) 650-6500 • (916) 650-6650 FAX
 Internet Address: www.cdph.ca.gov

