



Food and Drug Branch – Device Recalls

CALIFORNIA DEVICE RECALL INFORMATION SHEET

Mesa Biotech recalls Accula SARS-CoV-2 Test for Increased False Positives

Recall Date	Product Description	Recalling Firm	Recall Reason
4/06/2022	Accula SARS-CoV-2 Test REF: COV4100	MESA BIOTECH INC San Diego, California	SARS-CoV-2 Test has an increased potential for false positive results, which may lead to a delay in correct diagnosis and treatment.

Recall Class	Product Identification	Distribution	Affected Dates
I	UDI/DI: B540COV41000; P22006-026 - P22032-011	Nationwide including California	December 2021 – February 2022

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#)

CDPH Food and Drug Branch
 MS 7602 • P.O. Box 997435 • Sacramento, CA 95899-7435
 (916) 650-6500 • (916) 650-6650 FAX
 Internet Address: www.cdph.ca.gov

