



CALIFORNIA DEVICE RECALL INFORMATION SHEET

GE Healthcare Recalls MRI Software due to Slicing Error

Recall Date	Product Description	Recalling Firm	Recall Reason
2/22/2023	EVIS EXERA III Colonovideoscope For endoscopy and endoscopic surgery within the lower digestive tract (including the anus, rectum, sigmoid colon, colon, and ileocecal valve) Model/Serial: CF-Q180AL	Olympus Corporation of the Americas Center Valley, Pennsylvania	A single CF-Q180AL colonovideoscope was utilized in a veterinary endoscopy procedure in advance of being assigned to a medical facility as a service loaner in error, potential for microbial contamination

Recall Class	Product Identification	Distribution	Affected Dates
II	EVIS EXERA III UDI: 04953170202315 Serial number 2807443	1 unit in California	December 2022 and prior

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#)

