



Food and Drug Branch – Device Recalls

CALIFORNIA DEVICE RECALL INFORMATION SHEET

Philips Respironics Recalls Philips Respironics V60/V60 Plus Ventilator For Power Failure

Recall Date	Product Description	Recalling Firm	Recall Reason
10/4/2023	Philips Respironics V60/V60 Plus Ventilator, PCBA Part Number: 453561544461 V60 Device Part Numbers 1053617	Philips Respironics Murrysville, Pennsylvania	Power Management PCBA's may malfunction, causing a power failure leading to ventilator loss of function.

Recall Class	Product Identification	Distribution	Affected Dates
I	Philips Respironics V60 Device Serial # PCBA Product # PCBA Serial # 201019564 453561544461 SC23120K8 201014394 453561544461 ...	8 Units Nationwide including California	August 2023 and prior

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#)

