



Food and Drug Branch – Device Recalls

CALIFORNIA DEVICE RECALL INFORMATION SHEET

Physio-Control, Inc. Recalls AC Power Adapter For Monitor/Defibrillator Batteries Charging Problem For Monitor/Defibrillator (ACPA Connected) Power Up Problem

Recall Date	Product Description	Recalling Firm	Recall Reason
12/27/2023	Ac Power Adapter For use with the LIFEPAK 15 Monitor/Defibrillator, REF: 11140-000098	PHYSIO-CONTROL, INC. Redmond, Washington	AC Power Adapter (ACPA) may not charge the monitor/defibrillator batteries and may not allow the monitor/defibrillator to power on while plugged into the ACPA.

Recall Class	Product Identification	Distribution	Affected Dates
II	Ac Power Adapter UDI-DI: 00883873940204, Adapters manufactured between 11/3/2021 - 4/24/2023	225 Units in California	December, 2023 and prior

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [AC Power Adapter](#)

