



CALIFORNIA DEVICE RECALL INFORMATION SHEET

Preat Corp Is Recalling Neodent Healing Abutment For Possible Microleakage

Recall Date	Product Description	Recalling Firm	Recall Reason
11/1/2023	Neodent GM X 2mm Healing Abutment Preat REF 9007121, Rx Only, Non-Sterile, Single Use Only. Dental	PREAT CORP Santa Maria, California	The healing abutments contain a dimensional condition that may lead to a gap between the healing abutment and the implant allowing for microleakage
11/1/2023	Neodent GM X 4mm Healing Abutment Preat REF 9007123, Rx Only, Non-Sterile, Single Use Only. Dental	PREAT CORP Santa Maria, California	The healing abutments contain a dimensional condition that may lead to a gap between the healing abutment and the implant allowing for microleakage
11/1/2023	Neodent GM X 4mm Healing Abutment (Wide) Preat REF 9007124, Rx Only, Non-Sterile, Single Use Only. Dental	PREAT CORP Santa Maria, California	The healing abutments contain a dimensional condition that may lead to a gap between the healing abutment and the implant allowing for microleakage

Recall Class	Product Identification	Distribution	Affected Dates
II	Neodent GM X 2mm Healing Abutment UDI-DI: 00842092161364/Lot # OF018356	13 Units Nationwide including California	18 July 2023 and prior
II	Neodent GM X 4mm Healing Abutment UDI-DI: 00842092161388/Lot #OF018357	23 Units Nationwide including California	18 July 2023 and prior
II	Neodent GM x 4mm Healing Abutment (Wide) UDI-DI: 00842092161395/Lot #OF018355	4 Units Nationwide including California	18 July 2023 and prior

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#)