



Food and Drug Branch – Device Recalls

CALIFORNIA DEVICE RECALL INFORMATION SHEET

Remington Medical Recalls Automatic Cutting Needle for Labeling Error

Recall Date	Product Description	Recalling Firm	Recall Reason
1/04/2023	Remington MEDICAL, Automatic Cutting Needle 18ga (1.3mm) x 20cm, REF NAC-1820M (1)	Remington Medical, Inc. Alpharetta, Georgia	The stylet is longer than specified on the label

Recall Class	Product Identification	Distribution	Affected Dates
II	UDI/DI 00813079020936, Lot Number 2224920, exp. 2025-09/06	110 units in California	December 2022 – September 2025

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#)

