



Food and Drug Branch – Device Recalls

CALIFORNIA DEVICE RECALL INFORMATION SHEET

Sam Chest Seal Recalled for Packaging Error

Recall Date	Product Description	Recalling Firm	Recall Reason
01/28/2022	SAM CHEST SEAL REF CS203-EN Combo LOT Y060321-09	Sam Medical Products Tualatin, Oregon	Due to partially assembled valved dressing that has a hole within it instead of the intended non-valved dressing in the package resulting in that the dressing can not be centered over the open chest wound and create an occlusive seal due to the hole being present.

Recall	Product Identification	Distribution	Affected Dates
II	Model Number: CS203-EN UDI Code: (01)10822045000190(10) Y060321-09(17)270601 (Case Carton) UDI Code: (01)1 0822 045000193(10)Y060321-09(17) 270601 (Each Pouch/Package) Lot Number: Y060321-09	2352 Nationwide including California	September 2021 – November 2021

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#)

CDPH Food and Drug Branch
 MS 7602 • P.O. Box 997435 • Sacramento, CA 95899-7435
 (916) 650-6500 • (916) 650-6650 FAX
 Internet Address: www.cdph.ca.gov

