

State of California—Health and Human Services Agency

California Department of Public Health



Food and Drug Branch - Device Recalls

CALIFORNIA DEVICE RECALL INFORMATION SHEET

Smiths Medical ASD Recalls Devices Employing Fast Flow Fluid Warming and Irrigation System Disposables for leaching of Aluminum from Heat Excahangers

Recall Date	Product Description	Recalling Firm	Recall Reason
8/17/2021	Level 1 Fluid Warmer Model H-1000, H-500, H-1025	Smiths Medical ASD Inc. Minneapolis, Minnesota	Fast Flow Fluid Warming and Irrigation System Disposables, contain aluminum heat exchangers which have the potential to leach high levels of aluminum during the infusion of blood products, were distributed.
8/17/2021	Level 1 Fluid Warming System Model H-1028, H-1200	Smiths Medical ASD Inc. Minneapolis, Minnesota	Aluminum leaching
8/17/2021	Level 1 Normothermic I.V. Fluid Administration Set Model D-100, D-300, D-50, D- 60HL, D-70, DI-100, DI-300, DI-50, DI-60HL, DI-70	Smiths Medical ASD Inc. Minneapolis, Minnesota	Aluminum leaching

Recall Class	Product Identification	Distribution	Affected Dates
I	Level 1 Fluid Warmer H1000	268 Devices in California	August 2021 and prior
I	Level 1 Fluid Warmer H-1000-INT-230	1 Device in California	August 2021 and prior
I	Level 1 Fluid Warmer H-500	12 Devices in California	August 2021 and prior

I	Level 1 Normothermic I.V. Fluid Administration Set Model D-100	51,221 Devices in California	August 2021 and prior
I	Level 1 Normothermic I.V. Fluid Administration Set Model D-300	4,900 Devices in California	August 2021 and prior
I	Level 1 Normothermic I.V. Fluid Administration Set Model D-70	49,020 Devices in California	August 2021 and prior
I	Level 1 Normothermic I.V. Fluid Administration Set Model DI-70	200 Devices in California	August 2021 and prior
I	Level 1 Normothermic I.V. Fluid Administration Set Model D-60HL	3,750 Devices in California	August 2021 and prior
I	NORMOFLO Irrigation Fluid Warmer Model H-1129	46 Devices in California	August 2021 and prior

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE <u>FDA WEBSITE</u>

