



Food and Drug Branch – Device Recalls

**CALIFORNIA DEVICE RECALL INFORMATION SHEET**

**Teleflex Endotracheal Tubes Models: Slick Set, Flexi-Set, and Murphy for Disconnection of the Connector**

Recall Date	Product Description	Recalling Firm	Recall Reason
7/5/2023	<b>Slick Set Cuffed</b> REF 170050, 170055, 170060, 170065, 170070, 170075, 170080, 170085, 170090, 170095	<b>TELEFLEX LLC</b> Morrisville, NC	Reports of disconnection of the 15mm connector from the endotracheal tube (ET tube) for the affected products.
7/5/2023	<b>Slick Set Uncuffed</b> REF 170120, 170125, 170130, 170135, 170140, 170145, 170150, 170155	<b>TELEFLEX LLC</b> Morrisville, NC	Reports of disconnection.
7/5/2023	<b>Flexi-Set Cuffed</b> REF 504550, 504555, 504560, 504565, 504570, 504575, 504580, 504585, 504590	<b>TELEFLEX LLC</b> Morrisville, NC	Reports of disconnection.
7/5/2023	<b>Flexi-Set Uncuffed</b> REF 506525, 506530, 506535, 506540, 506545, 506550	<b>TELEFLEX LLC</b> Morrisville, NC	Reports of disconnection.
7/5/2023	<b>Murphy Uncuffed</b> REF 100180030, 100180035, 100180040, 100180045, 100180050, 100180055, 100180060, 100180065, 100180070, 100181030, 100181040, 100181045, 100181050, 100181055, 100181060, 100181065, 100382025, 100382030, 100382035, 100382040,	<b>TELEFLEX LLC</b> Morrisville, NC	Reports of disconnection.

	100382045, 100382050, 100382055, 100382060, 100382065		
7/5/2023	<b>Murphy--Low Pressure Cuff</b> REF 111780035, 111780040, 111780045, 111780050, 111780055, 111780060, 111780065, 111780070, 111780075, 111780080, 111780085, 111780090, 111781035, 111781040, 111781045, 111781050, 111781055, 111781060, 111781065, 111781070, 111781075, 111781080	<b>TELEFLEX LLC</b> Morrisville, NC	Reports of disconnection.

Recall Class	Product Identification	Distribution	Affected Dates
I	<b>Slick Set Cuffed</b> UDI/DI: 4026704319534; Batch numbers: 18FG05-KME22H0342...	181,635,049 Units Nationwide	May 2023 and Prior
I	<b>Slick Set Uncuffed</b> UDI/DI 14026704644879, Batch Numbers: 18JG23-KME22J0262...	97,342 Units Nationwide	May 2023 and Prior
I	<b>Flexi-Set Cuffed</b> UDI/DI 4026704547579, Batch Numbers: 18FG03-20CT44...	2,557,072 Units Nationwide	May 2023 and Prior
I	<b>Flexi-Set Uncuffed</b> UDI/DI 4026704547661, Batch Numbers: 18FG05-20BT51...	107,857 Units Nationwide	May 2023 and Prior
I	<b>Murphy Uncuffed</b> UDI/DI 14026704340436, Batch Numbers: 18GG26-KME22K2262	189,606 Units Nationwide	May 2023 and Prior
I	<b>Murphy--Low Pressure Cuff</b> UDI/DI 14026704341037, Batch Numbers: 18GG0-KME22B0923...	1,701,536 Units Nationwide	May 2023 and Prior

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#)

---

CDPH Food and Drug Branch  
MS 7602 ● P.O. Box 997435 ● Sacramento, CA 95899-7435  
(916) 650-6500 ● (916) 650-6650 FAX  
Internet Address: [www.cdph.ca.gov](http://www.cdph.ca.gov)

