



Food and Drug Branch – Device Recalls

**CALIFORNIA DEVICE RECALL INFORMATION SHEET**

**Tornier Recalls Stryker Tornier Augmented Glenoid Lateralized Baseplate For Manufacturing With 6mm Spacers**

Recall Date	Product Description	Recalling Firm	Recall Reason
10/25/2023	<b>Tornier</b> stryker Tornier Perform Reversed Augmented Glenoid Lateralized Baseplate, 25mm, Offset +3mm, REF DWJ502	<b>TORNIER,INC</b> Bloomington, Minnesota	One lot of Tournier Perform Reversed Lateralized Baseplates were found to have been manufactured with 6mm spacers

Recall Class	Product Identification	Distribution	Affected Dates
II	<b>Tornier</b> UDI/DI 10846832062017, Lot Number 1756123	30 Units Nationwide including California	September 2023 and prior

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#)

