



Food and Drug Branch – Device Recalls

CALIFORNIA DEVICE RECALL INFORMATION SHEET

Zimmer Recalls M/L Taper Hip Prosthesis Extended Offset For Size Irregularity

Recall Date	Product Description	Recalling Firm	Recall Reason
9/13/2023	M/L Taper Hip Prosthesis Extended Offset Reduced Neck Length, Size 4 (Item Number 00771100640) and Size 6 (Item Number 00771100440)	ZIMMER, INC. Warsaw, Indiana	The outer package labeling and product etch are for a Size 6, however, the implant is a Size 4, and vice versa.

Recall Class	Product Identification	Distribution	Affected Dates
II	M/L Taper Hip Prosthesis Extended Offset Item Number: 00771100640, UDI-DI: 00889024131699, Lot Number: 65236202; Item Number: 00771100440, UDI-DI: 00889024131613, Lot Number: 65249215	24 Units Nationwide including California	26 July 2023, and prior

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](https://www.fda.gov)

